

Declaration of Conformity

(in accordance with ISO/IEC 17050-1)

Company Information:

American Orthodontics
3524 Washington Avenue
Sheboygan, WI 53081

Certificate Number: 201706 -08C

Certificate Issue Date: 06/12/17
Certificate Expiry Date: 12/15/17
This certificate is valid until Expiry Date listed above.

Medical Device Detail:

Product/Family Category:
Lingual Attachments

Classification: *Classification is according to Directive 93/42/EEC, Annex IX*
Class IIa *per Rule 5, 3rd indent*

Product(s) Tradename(s):
Ceramic Button

CE Authorized Representative:

Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

Assessment Details:

Notified Body: UL International (UK) Ltd.
CE Certificate #: 860.170201
Issue Date: 02/01/17
Route of Directive: Annex V, Sect. 3
of the Directive 93/42/EEC on Medical Devices.

Standards Applied:

Harmonized Standards: EN ISO 13485, IEC 62336-1, EN ISO 14971, EN 1041, EN ISO 10993-1, 10993-3, 10993-5, 10993-11, EN 980

Other Standards: ISO 10993-10, ISO 7405, ISO 10271, ISO 15223-1, ISO 15223-2, ISO 9333

Declaration statement:

Each kind of medical device to which the technical documentation applies complies with the applicable provisions of the essential principles and the classifications rules before being shipped. The above products are in compliance with Annex I Essential Requirements of Directive 93/42/EEC. The product is CE marked.

Signed for and on behalf of:

American Orthodontics
Sheboygan, WI

Company Representative:


Trang Adams / Regulatory Affairs Specialist

